Survey:

BI4/23

month:

October 2023

Participant-No:

0004965

date of issue:

14.10.23



Referenzinstitut für Bioanalytik

diaglobal GmbH Dr. Rainer Schlegel Haus 41 Köpenicker Str. 325 12555 Berlin



Survey directors

Prof. Dr. C. Knabbe Prof. Dr. Dr. K.P. Kohse Prof. Dr. M. Wearmaiei

> Head of RfB Dr. A. Kessler

Bonn, 22. November 2023

Certificate

We confirm that you have participated in the survey for neonatal bilirubin in serum.

You have met the requirements of the survey for the following analytes:

Total bilirubin

(2)

This certificate is according to the current version of the RiliBÄK valid until the end of April 2024.

The number in parentheses characterizes the analytical method used.

The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

http://www.rfb.bio/ogi/certificate?tin=0004965&rv_Id=B1234 &version=0.8.urid=01294.46810A49488A396B31B73510240 Survey:

BI4/23

month:

October 2023

Participant-No: 0004965



Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed. If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T)	Certification:	+ = fulfilled (quotient D/Dmax <= 1.0)
	Dmax= maximum allowable amount of difference in measurement, partly given by		- = not fulfilled (quotient D/Dmax > 1.0
	annax 1 of the guideline of the BÄK (Dt. Ärzteblatt 111, Heft 38, 19.9.2014).		± = certification cancelled because of technical
1	T = target value, either reference method value or assigned value 11 1111 = lower resp. upper limit		and/or analytical reasons

	С	M		R	D/Dmax	Т	LL	UL		_1			
Total bilirubin [mg/dl]	+	2	A B	9.46 12.6	0.45 0.10	8.60 12.3	6.70 9.59	10.5 15.1					
]				ı	,	