

Survey: **KS4/23**
month: **May 2023**
Participant-No: **0004965**
date of issue: **13.05.23**



Referenzinstitut für Bioanalytik

diaglobal GmbH
Dr. Rainer Schlegel
Haus 41
Köpenicker Str. 325
12555 Berlin



Survey directors

Prof. Dr. C. Knabbe
Prof. Dr. Dr. K.P. Kohse
Prof. Dr. M. Neumaier

Head of RfB

Dr. A. Kessler

Bonn, 23. June 2023

Certificate

We confirm that you have participated in the survey for clinical chemical analytes in serum.

You have met the requirements of the survey for the following analytes:

cholesterol	(1)	glucose	(3)	lactate	(2)
triglycerides	(2)	CK	(1)	AST	(1)
ALT	(1)				

This certificate is according to the current version of the RiliBÄK valid until the end of November 2023.

The number in parentheses characterizes the analytical method used.
The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.

http://www.rfb.bio/sg/identif/cafes/?form=0004965&ov_d=KS234&version=0&uid=ED388770413403D04AE394EE549FD1

Survey: **KS4/23**
 month: **May 2023**
 Participant-No: **0004965**

Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed.
 If all analytes are listed on the certificate no participation certificate is printed.

Legend:	C=Certification, M=No of method, R=your result, D=difference (R-T) Dmax= maximum allowable amount of difference in measurement, partly given by annex 1 of the guideline of the BÄK (Dt. Ärzteblatt 111, Heft 38, 19.9.2014). T = target value, either reference method value or assigned value, LL UL = lower resp. upper limit	Certification:	+ = fulfilled (quotient D/Dmax <= 1.0) - = not fulfilled (quotient D/Dmax > 1.0) ± = certification cancelled because of technical and/or analytical reasons
---------	--	----------------	---

	C	M	R	D/Dmax	T	LL	UL							
cholesterol [mg/dl]	+	1	A	219	-0.33	229	199	259						
			B	170	-0.13	173	150	196						
glucose [mg/dl]	+	3	A	111	-0.63	123	104	142						
			B	78.8	-0.41	84.0	71.4	96.6						
lactate [mmol/l]	+	2	A	4.62	-0.01	4.63	3.78	5.48						
			B	4.22	0.35	3.96	3.24	4.69						
triglycerides [mg/dl]	+	2	A	134	-0.09	136	114	158						
			B	200	0.46	186	156	216						
CK [U/l]	+	1	A	250	0.20	240	192	288						
			B	394	0.47	360	288	432						
AST [U/l]	+	1	A	251	-0.47	279	220	338						
			B	87.8	-0.03	88.4	69.8	107						
ALT [U/l]	+	1	A	204	-0.33	219	173	265						
			B	102	0	102	80.5	124						