Survey:

HA2/14

month:

May 2014

Participant-No:

0004965

date of issue:

diaglobal GmbH Dr. Rainer Schlegel Martina Schmidt Köpenicker Str. 325

12555 Berlin

14.05.14



Referenzinstitut für Bioanalytik



Ringversuchsleiter

Prof. Dr. C. Knabbe Prof. Dr. Dr. K.P. Kohse Prof. Dr. M. Neumaier

Leitung RfB

Dr. R. Kruse Dr. W. J. Geilenkeuser

Bonn, 5. June 2014

Certificate

We confirm that you have participated in the survey for full blood count.

You have met the requirements of the survey for the following analytes:

hemoglobin

(4)

erythrocytes

(7)

hematocrit

(7)

This certificate is valid until the end of November 2014.

hear, p. hola

Lagana

M.J. Jilulum

The number in parentheses characterizes the analytical method used. The assignment of the number to the respective method and/or the respective instrument is to be taken from the total evaluation.



HA2/14

month:

May 2014

Participant-No:

0004965



Listing and Evaluation of all your results

Explanations

Certificate

A certificate is issued (given) for an analyte only if the basis for an evaluation of the accuracy is given by the guidelines of the German Medical Association and/or if an evaluation is possible in analogy to these guidelines (see comments on the evaluation) and, if both results for an analyte are within the given acceptance limits. (marked as '+' below C)

Certificate of participation

In the participation certificate all analytes which are included in the list on this page are listed. If all analytes are listed on the certificate no participation certificate is printed.

Legend: C=Certification, M=No of method, R=your result, D=difference (R-T)

Dmax= maximum allowable amount of difference in measurement, partly given by

annax 1 of the guideline of the BÄK (Dt. Ärzteblatt 105, Heft 7, 15.2.2008).

T = target value, either reference method value or assigned value, LL | UL = lower resp. upper limit

+ = fulfilled (quotient | D/Dmax | <= 1.0)
- = not fulfilled (quotient | D/Dmax | > 1.0)
- = not fulfilled (quotient | D/Dmax | > 1.0)
- = certification cancelled because of technical and/or analytical reasons

	С	M		R	D/Dmax	Т	LL	UL				
hemoglobin [g/dl] erythrocytes [10e6/ul] hematocrit [%]	+	7	A B	15.9 14.2 5.12 4.56 44.3 39.6	0.30 0.44 -0.10 -0.16 0.30 0.34	15.6 13.8 5.16 4.62 43.1 38.4	14.6 12.9 4.74 4.25 39.2 34.9	16.6 14.7 5.58 4.99 47.0 41.9				